

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073295.D
 Acq On : 01 Jul 2022 14:22
 Operator : JC\MD
 Sample : N3485-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-20220623

Quant Time: Jul 04 05:04:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	303263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	502930	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	466162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	234873	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	188015	42.946	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.900%
35) Dibromofluoromethane	7.951	113	156930	46.900	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.800%
50) Toluene-d8	10.381	98	528547	44.475	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.960%
62) 4-Bromofluorobenzene	12.669	95	240413	52.641	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.280%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.293	59	6446	6.078	ug/l #	84
16) Acetone	4.234	43	19629	7.988	ug/l	94
25) 2-Butanone	7.281	43	5309	1.357	ug/l	96
52) Toluene	10.433	92	2902	0.296	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073295.D
Acq On : 01 Jul 2022 14:22
Operator : JC\MD
Sample : N3485-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-20220623

Quant Time: Jul 04 05:04:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

