

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073296.D
 Acq On : 01 Jul 2022 14:47
 Operator : JC\MD
 Sample : N3485-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01-20220623

Quant Time: Jul 04 05:05:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	301643	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	491362	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	448269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	208231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	186794	42.896	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.800%
35) Dibromofluoromethane	7.957	113	156162	47.769	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	10.380	98	517160	44.542	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.080%
62) 4-Bromofluorobenzene	12.669	95	221354	49.609	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.220%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.293	59	12785	12.120	ug/l #	86
16) Acetone	4.240	43	18543	7.586	ug/l	96
25) 2-Butanone	7.281	43	5407	1.389	ug/l	93
52) Toluene	10.445	92	4088	0.426	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073296.D
Acq On : 01 Jul 2022 14:47
Operator : JC\MD
Sample : N3485-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01-20220623

Quant Time: Jul 04 05:05:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

