

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073297.D
 Acq On : 01 Jul 2022 15:12
 Operator : JC\MD
 Sample : N3569-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-05-20

Quant Time: Jul 04 05:05:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	299104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	491169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	448375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	223226	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	186281	43.142	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.280%		
35) Dibromofluoromethane	7.951	113	157523	48.204	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.400%		
50) Toluene-d8	10.380	98	518860	44.705	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.420%		
62) 4-Bromofluorobenzene	12.668	95	231974	52.010	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.020%		
Target Compounds						Qvalue
4) Vinyl Chloride	2.398	62	1438	0.399	ug/l #	88
16) Acetone	4.234	43	10143	4.185	ug/l #	73
27) cis-1,2-Dichloroethene	7.251	96	8578	1.925	ug/l	93
39) Methylcyclohexane	9.392	83	2359	0.377	ug/l #	84
52) Toluene	10.451	92	2584	0.269	ug/l	97
73) Isopropylbenzene	12.521	105	12073	0.626	ug/l	99
78) n-propylbenzene	12.857	91	16561	0.773	ug/l	97
83) tert-Butylbenzene	13.262	119	3404	0.242	ug/l	89
84) 1,2,4-Trimethylbenzene	13.310	105	41630	2.654	ug/l	99
85) sec-Butylbenzene	13.439	105	6076	0.321	ug/l	97
87) 1,3-Dichlorobenzene	13.551	146	1969	0.245	ug/l	82
88) 1,4-Dichlorobenzene	13.627	146	21295	2.643	ug/l	89
95) Naphthalene	15.433	128	17021	0.989	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073297.D
Acq On : 01 Jul 2022 15:12
Operator : JC\MD
Sample : N3569-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-05-20

Quant Time: Jul 04 05:05:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

