

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073298.D
 Acq On : 01 Jul 2022 15:36
 Operator : JC\MD
 Sample : N3569-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-05-30

Quant Time: Jul 04 05:05:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	311267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	511055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	476142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	238293	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	191824	42.689	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.380%
35) Dibromofluoromethane	7.957	113	164531	48.390	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.780%
50) Toluene-d8	10.380	98	543306	44.990	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.980%
62) 4-Bromofluorobenzene	12.669	95	244785	52.746	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.500%
Target Compounds						
					Qvalue	
16) Acetone	4.246	43	4729	1.875	ug/l #	86
39) Methylcyclohexane	9.392	83	2214	0.340	ug/l #	52
85) sec-Butylbenzene	13.439	105	24523	1.215	ug/l	85
89) n-Butylbenzene	13.880	91	5114	0.376	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073298.D
Acq On : 01 Jul 2022 15:36
Operator : JC\MD
Sample : N3569-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-05-30

Quant Time: Jul 04 05:05:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

