

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073301.D
 Acq On : 01 Jul 2022 16:50
 Operator : JC\MD
 Sample : N3569-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-05-60

Quant Time: Jul 04 05:06:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	291023	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	487488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	441843	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	206274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	181425	43.184	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.360%		
35) Dibromofluoromethane	7.951	113	154414	47.610	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.220%		
50) Toluene-d8	10.381	98	506232	43.947	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	87.900%		
62) 4-Bromofluorobenzene	12.669	95	217097	49.042	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.080%		
Target Compounds						
44) Trichloroethene	9.157	130	1748	0.446	ug/l	72
64) Tetrachloroethene	10.916	164	979	0.271	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073301.D
Acq On : 01 Jul 2022 16:50
Operator : JC\MD
Sample : N3569-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-05-60

Quant Time: Jul 04 05:06:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

