

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073303.D
 Acq On : 01 Jul 2022 17:39
 Operator : JC\MD
 Sample : N3569-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-05-80

Quant Time: Jul 04 05:07:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	303302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	497962	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	452615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	221850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	189900	43.371	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.740%
35) Dibromofluoromethane	7.951	113	159458	48.131	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.260%
50) Toluene-d8	10.380	98	522030	44.365	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.740%
62) 4-Bromofluorobenzene	12.668	95	229357	50.721	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.440%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	3942	1.604	ug/l	93
30) Chloroform	7.745	83	28653	3.917	ug/l	99
32) 1,1,1-Trichloroethane	7.939	97	8740	1.341	ug/l #	50
38) Carbon Tetrachloride	8.133	117	13820	2.622	ug/l #	95
44) Trichloroethene	9.151	130	25306	6.321	ug/l	91
64) Tetrachloroethene	10.922	164	9478494	2564.574	ug/l	87
90) Hexachloroethane	14.145	117	90602	32.015	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073303.D
Acq On : 01 Jul 2022 17:39
Operator : JC\MD
Sample : N3569-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-05-80

Quant Time: Jul 04 05:07:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

