

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073304.D
 Acq On : 01 Jul 2022 18:04
 Operator : JC\MD
 Sample : N3569-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-SPC-95-80

Quant Time: Jul 04 05:07:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	289576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	484277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	440268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	214664	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	182732	43.712	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.420%
35) Dibromofluoromethane	7.957	113	153210	47.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	10.380	98	511354	44.686	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.380%
62) 4-Bromofluorobenzene	12.668	95	219199	49.845	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.680%
Target Compounds						
					Qvalue	
16) Acetone	4.245	43	3840	1.636	ug/l #	86
30) Chloroform	7.751	83	27773	3.976	ug/l	94
32) 1,1,1-Trichloroethane	7.939	97	8680	1.395	ug/l	95
38) Carbon Tetrachloride	8.133	117	15123	2.950	ug/l	95
44) Trichloroethene	9.151	130	25618	6.580	ug/l	97
64) Tetrachloroethene	10.921	164	9737427	2708.519	ug/l #	87
90) Hexachloroethane	14.145	117	113068	41.292	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073304.D
Acq On : 01 Jul 2022 18:04
Operator : JC\MD
Sample : N3569-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-SPC-95-80

Quant Time: Jul 04 05:07:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

