

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
 Data File : VN073282.D
 Acq On : 01 Jul 2022 08:53
 Operator : JC\MD
 Sample : VN0701MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0701MBL01

Quant Time: Jul 01 16:15:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	295442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	482809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	444539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	208948	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	184933	43.360	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.720%
35) Dibromofluoromethane	7.945	113	153003	47.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.260%
50) Toluene-d8	10.375	98	504510	44.222	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.440%
62) 4-Bromofluorobenzene	12.669	95	220897	50.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.760%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070122\
Data File : VN073282.D
Acq On : 01 Jul 2022 08:53
Operator : JC\MD
Sample : VN0701MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0701MBL01

Quant Time: Jul 01 16:15:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.016 min Scan# 1042
Delta R.T. -0.000 min
Lab File: VN073282.D
Acq: 01 Jul 2022 08:53




Instrument :
MSVOA_N
ClientSampleId :
VN0701MBL01

Tgt Ion:168 Resp: 295442
Ion Ratio Lower Upper
168 100
99 62.9 42.3 63.5



#33
1,2-Dichloroethane-d4
Concen: 43.360 ug/l
RT: 8.369 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VN073282.D
Acq: 01 Jul 2022 08:53

Tgt Ion: 65 Resp: 184933
Ion Ratio Lower Upper
65 100
67 51.1 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.898 min Scan# 1192

Delta R.T. 0.000 min

Lab File: VN073282.D

Acq: 01 Jul 2022 08:53

Instrument :

MSVOA_N

ClientSampleId :

VN0701MBL01




Tgt Ion:114 Resp: 482809

Ion Ratio Lower Upper

114 100

63 21.4 0.0 40.0

88 15.5 0.0 32.6



#35

Dibromofluoromethane

Concen: 47.632 ug/l

RT: 7.945 min Scan# 1030

Delta R.T. -0.006 min

Lab File: VN073282.D

Acq: 01 Jul 2022 08:53




Tgt Ion:113 Resp: 153003


Ion Ratio Lower Upper

113 100

111 102.5 81.4 122.0

192 18.5 17.0 25.6








#50
Toluene-d8
Concen: 44.222 ug/l
RT: 10.375 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN073282.D
Acq: 01 Jul 2022 08:53




Instrument :
MSVOA_N
ClientSampleId :
VN0701MBL01


Tgt Ion: 98 Resp: 504510
Ion Ratio Lower Upper
98 100
100 65.6 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 50.384 ug/l
RT: 12.669 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN073282.D
Acq: 01 Jul 2022 08:53

Tgt Ion: 95 Resp: 220897
Ion Ratio Lower Upper
95 100
174 74.3 0.0 167.6
176 72.4 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.680 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN073282.D
Acq: 01 Jul 2022 08:53

Instrument :
MSVOA_N
ClientSampleId :
VN0701MBL01

Tgt Ion:117 Resp: 444539
Ion Ratio Lower Upper
117 100
82 57.6 44.2 66.2
119 32.0 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.610 min Scan# 1993
Delta R.T. -0.000 min
Lab File: VN073282.D
Acq: 01 Jul 2022 08:53

Tgt Ion:152 Resp: 208948
Ion Ratio Lower Upper
152 100
115 61.2 28.9 86.7
150 156.2 0.0 356.2

