

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
 Data File : VN082767.D
 Acq On : 01 Jul 2024 15:52
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 02 00:09:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	102190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	194600	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	191600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	74863	47.504	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.000%
35) Dibromofluoromethane	8.171	113	63343	49.439	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	10.565	98	253368	52.878	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.760%
62) 4-Bromofluorobenzene	12.853	95	94741	57.576	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	115.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
Data File : VN082767.D
Acq On : 01 Jul 2024 15:52
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 02 00:09:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

