

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
 Data File : VN082768.D
 Acq On : 01 Jul 2024 16:16
 Operator : JC\MD
 Sample : PB161815TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161815TB

Quant Time: Jul 02 00:09:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	106456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	205452	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	208488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	80315	48.922	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.840%		
35) Dibromofluoromethane	8.171	113	67034	49.556	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.120%		
50) Toluene-d8	10.570	98	269209	53.217	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	106.440%		
62) 4-Bromofluorobenzene	12.847	95	100677	57.951	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery =	115.900%		
Target Compounds					Qvalue	
16) Acetone	4.436	43	18522	29.778	ug/l #	78
20) Methylene Chloride	5.288	84	9704	6.165	ug/l #	74
43) Isopropyl Acetate	8.694	43	108250	29.704	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
Data File : VN082768.D
Acq On : 01 Jul 2024 16:16
Operator : JC\MD
Sample : PB161815TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161815TB

Quant Time: Jul 02 00:09:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

