

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
 Data File : VN082769.D
 Acq On : 01 Jul 2024 16:40
 Operator : JC\MD
 Sample : PB161815ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161815ZHE#01

Quant Time: Jul 02 00:09:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	107177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	210979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	216717	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	81286	49.180	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.360%
35) Dibromofluoromethane	8.171	113	69609	50.112	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.220%
50) Toluene-d8	10.571	98	275519	53.037	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.080%
62) 4-Bromofluorobenzene	12.853	95	101128	56.686	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	113.380%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	19534	31.194	ug/l	92
20) Methylene Chloride	5.277	84	10102	6.375	ug/l	87
43) Isopropyl Acetate	8.694	43	142105	37.897	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
Data File : VN082769.D
Acq On : 01 Jul 2024 16:40
Operator : JC\MD
Sample : PB161815ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161815ZHE#01

Quant Time: Jul 02 00:09:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

