

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
 Data File : VN082770.D
 Acq On : 01 Jul 2024 17:05
 Operator : JC\MD
 Sample : PB161815ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161815ZHE#02

Quant Time: Jul 02 00:10:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	117669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	222151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	205666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	84921	46.798	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.600%
35) Dibromofluoromethane	8.171	113	73591	50.314	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.620%
50) Toluene-d8	10.565	98	286753	52.424	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.840%
62) 4-Bromofluorobenzene	12.847	95	97513	51.911	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	103.820%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	21523	31.305	ug/l	94
20) Methylene Chloride	5.277	84	11859	6.817	ug/l #	75
43) Isopropyl Acetate	8.694	43	160487	40.606	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
Data File : VN082770.D
Acq On : 01 Jul 2024 17:05
Operator : JC\MD
Sample : PB161815ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161815ZHE#02

Quant Time: Jul 02 00:10:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

