

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
 Data File : VN082772.D
 Acq On : 01 Jul 2024 17:53
 Operator : JC\MD
 Sample : PB161815ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161815ZHE#04

Quant Time: Jul 02 00:11:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	116288	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	219698	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	222416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	88743	49.485	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.960%
35) Dibromofluoromethane	8.165	113	73295	50.671	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.340%
50) Toluene-d8	10.565	98	280532	51.859	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.720%
62) 4-Bromofluorobenzene	12.847	95	106378	57.262	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	114.520%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	38339	56.426	ug/l	100
20) Methylene Chloride	5.283	84	7043	4.096	ug/l #	71
43) Isopropyl Acetate	8.694	43	175782	44.890	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
Data File : VN082772.D
Acq On : 01 Jul 2024 17:53
Operator : JC\MD
Sample : PB161815ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161815ZHE#04

Quant Time: Jul 02 00:11:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

