

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
 Data File : VN082773.D
 Acq On : 01 Jul 2024 18:18
 Operator : JC\MD
 Sample : PB161815ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161815ZHE#05

Quant Time: Jul 02 00:12:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	119148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	228482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	229439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	106592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	91164	49.615	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.220%
35) Dibromofluoromethane	8.171	113	75659	50.295	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.580%
50) Toluene-d8	10.571	98	293660	52.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.847	95	111738	57.835	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	115.680%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	37482	53.841	ug/l	98
20) Methylene Chloride	5.271	84	7589	4.308	ug/l #	75
43) Isopropyl Acetate	8.688	43	167374	41.166	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
Data File : VN082773.D
Acq On : 01 Jul 2024 18:18
Operator : JC\MD
Sample : PB161815ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161815ZHE#05

Quant Time: Jul 02 00:12:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

