

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
 Data File : VN082778.D
 Acq On : 01 Jul 2024 20:19
 Operator : JC\MD
 Sample : P3110-01DL 20X
 Misc : 1.50g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 163DL

Quant Time: Jul 02 00:14:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	114763	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	212949	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	217287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101742	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	84449	47.716	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 95.440%			
35) Dibromofluoromethane	8.165	113	70301	50.142	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.280%			
50) Toluene-d8	10.565	98	270140	51.521	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.040%			
62) 4-Bromofluorobenzene	12.847	95	106602	59.202	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 118.400%			
Target Compounds						Qvalue
39) Methylcyclohexane	9.606	83	6645	3.176	ug/l #	85
52) Toluene	10.630	92	194560	50.119	ug/l	98
67) Ethyl Benzene	11.965	91	143843	17.882	ug/l	100
68) m/p-Xylenes	12.071	106	212673	71.055	ug/l	97
69) o-Xylene	12.400	106	88766	30.774	ug/l	97
73) Isopropylbenzene	12.694	105	8786	1.184	ug/l	96
78) n-propylbenzene	13.035	91	20445	2.333	ug/l	84
80) 1,3,5-Trimethylbenzene	13.176	105	25064	4.112	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	114774	19.051	ug/l	97
89) n-Butylbenzene	14.059	91	6704	1.315	ug/l #	77
95) Naphthalene	15.641	128	47734	8.779	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070124\
Data File : VN082778.D
Acq On : 01 Jul 2024 20:19
Operator : JC\MD
Sample : P3110-01DL 20X
Misc : 1.50g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
163DL

Quant Time: Jul 02 00:14:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

