

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
 Data File : VN087249.D
 Acq On : 01 Jul 2025 09:59
 Operator : JC\MD
 Sample : VN0701MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0701MBL01

Quant Time: Jul 02 01:57:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	180690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	333432	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303022	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	148702	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	120086	52.954	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.900%
35) Dibromofluoromethane	8.177	113	108642	50.886	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.780%
50) Toluene-d8	10.571	98	399711	48.032	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.060%
62) 4-Bromofluorobenzene	12.847	95	134701	45.493	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
Data File : VN087249.D
Acq On : 01 Jul 2025 09:59
Operator : JC\MD
Sample : VN0701MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0701MBL01

Quant Time: Jul 02 01:57:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

