

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
 Data File : VN087250.D
 Acq On : 01 Jul 2025 10:20
 Operator : JC\MD
 Sample : VN0701WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0701WBL01

Quant Time: Jul 02 01:57:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	242463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	449705	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	413338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	208904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	162343	53.349	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.700%	
35) Dibromofluoromethane	8.171	113	144568	50.206	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.420%	
50) Toluene-d8	10.565	98	544006	48.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.940%	
62) 4-Bromofluorobenzene	12.847	95	188727	47.259	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 94.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
Data File : VN087250.D
Acq On : 01 Jul 2025 10:20
Operator : JC\MD
Sample : VN0701WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0701WBL01

Quant Time: Jul 02 01:57:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

