

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
 Data File : VN087254.D
 Acq On : 01 Jul 2025 11:57
 Operator : JC\MD
 Sample : Q2452-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Jul 02 02:00:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	218634	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	412667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	392142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	194168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	150894	54.991	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.980%	
35) Dibromofluoromethane	8.171	113	132430	50.119	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.240%	
50) Toluene-d8	10.565	98	507207	49.247	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.500%	
62) 4-Bromofluorobenzene	12.847	95	178523	48.716	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 97.440%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	20602	19.638	ug/l	100
18) Methyl Acetate	5.053	43	16183	6.738	ug/l	98
20) Methylene Chloride	5.300	84	12416	5.343	ug/l #	81
37) Ethyl Acetate	7.577	43	11476	2.971	ug/l	98
43) Isopropyl Acetate	8.694	43	7609	1.270	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
Data File : VN087254.D
Acq On : 01 Jul 2025 11:57
Operator : JC\MD
Sample : Q2452-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Jul 02 02:00:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

