

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
 Data File : VN087257.D
 Acq On : 01 Jul 2025 13:01
 Operator : JC\MD
 Sample : Q2469-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Jul 02 02:02:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	169443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	328621	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	323380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	153280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	125383	58.959	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 117.920%		
35) Dibromofluoromethane	8.171	113	108904	51.756	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.520%		
50) Toluene-d8	10.565	98	408962	49.863	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.720%		
62) 4-Bromofluorobenzene	12.847	95	143379	49.132	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 98.260%		
Target Compounds						
						Qvalue
16) Acetone	4.453	43	17757	21.840	ug/l	99
18) Methyl Acetate	5.053	43	15855	8.518	ug/l	98
20) Methylene Chloride	5.306	84	8449	4.691	ug/l	95
25) 2-Butanone	7.500	43	7778	6.122	ug/l	94
37) Ethyl Acetate	7.577	43	10923	3.551	ug/l	98
43) Isopropyl Acetate	8.700	43	6890	1.444	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070125\
Data File : VN087257.D
Acq On : 01 Jul 2025 13:01
Operator : JC\MD
Sample : Q2469-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Jul 02 02:02:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

