

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
 Data File : VN049674.D
 Acq On : 2 Jul 2018 17:19
 Operator : MD\SY
 Sample : PB110744 ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110744 ZHE#01

Quant Time: Jul 03 04:51:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1364130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2157564	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1794755	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	705698	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	915508	50.82	ug/l	0.00
Spiked Amount						
			Recovery	=	101.64%	
35) Dibromofluoromethane	7.59	113	751184	46.57	ug/l	0.00
Spiked Amount						
			Recovery	=	93.14%	
50) Toluene-d8	10.09	98	2899674	45.93	ug/l	0.00
Spiked Amount						
			Recovery	=	91.86%	
62) 4-Bromofluorobenzene	12.40	95	863896	40.08	ug/l	0.00
Spiked Amount						
			Recovery	=	80.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	54680	7.36	ug/l	95
18) Methyl Acetate	4.33	43	118631	2.23	ug/l	95
20) Methylene Chloride	4.55	84	33714	1.82	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
Data File : VN049674.D
Acq On : 2 Jul 2018 17:19
Operator : MD\SY
Sample : PB110744 ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110744 ZHE#01

Quant Time: Jul 03 04:51:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

