

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
 Data File : VN049676.D
 Acq On : 2 Jul 2018 18:11
 Operator : MD\SY
 Sample : PB110744 ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110744 ZHE#03

Quant Time: Jul 03 04:56:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1289271	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2087952	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1725347	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	687204	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	890115	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
35) Dibromofluoromethane	7.59	113	734198	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
50) Toluene-d8	10.09	98	2801510	45.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.70%	
62) 4-Bromofluorobenzene	12.41	95	822051	39.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.82%	
Target Compounds						
16) Acetone	3.83	43	54822	8.31	ug/l	95
18) Methyl Acetate	4.33	43	108563	1.97	ug/l	100
20) Methylene Chloride	4.55	84	32175	1.84	ug/l	95
25) 2-Butanone	6.86	43	10429	1.85	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070218\
Data File : VN049676.D
Acq On : 2 Jul 2018 18:11
Operator : MD\SY
Sample : PB110744 ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110744 ZHE#03

Quant Time: Jul 03 04:56:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

