

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
 Data File : VN049679.D
 Acq On : 2 Jul 2018 19:29
 Operator : MD\SY
 Sample : PB110744 ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110744 ZHE#06

Quant Time: Jul 03 05:02:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1267435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2061032	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1699502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	647748	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	869743	51.96	ug/l	0.00
Spiked Amount			Recovery	=	103.92%	
35) Dibromofluoromethane	7.59	113	728469	47.28	ug/l	0.00
Spiked Amount			Recovery	=	94.56%	
50) Toluene-d8	10.09	98	2748131	45.57	ug/l	0.00
Spiked Amount			Recovery	=	91.14%	
62) 4-Bromofluorobenzene	12.41	95	797968	38.75	ug/l	0.00
Spiked Amount			Recovery	=	77.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	51232	7.49	ug/l	90
18) Methyl Acetate	4.33	43	107407	2.02	ug/l	98
20) Methylene Chloride	4.55	84	19443	1.13	ug/l	86
25) 2-Butanone	6.85	43	6773	1.22	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
Data File : VN049679.D
Acq On : 2 Jul 2018 19:29
Operator : MD\SY
Sample : PB110744 ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110744 ZHE#06

Quant Time: Jul 03 05:02:46 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

