

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
 Data File : VN049680.D
 Acq On : 2 Jul 2018 19:55
 Operator : MD\SY
 Sample : PB110744 ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110744 ZHE#07

Quant Time: Jul 03 05:10:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1233085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2023133	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1674888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	671366	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	852275	52.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.68%	
35) Dibromofluoromethane	7.59	113	711845	47.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.12%	
50) Toluene-d8	10.09	98	2706411	45.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.42%	
62) 4-Bromofluorobenzene	12.40	95	782179	38.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.40%	

Target Compounds						Qvalue
16) Acetone	3.83	43	46191	6.34	ug/l	99
18) Methyl Acetate	4.33	43	111345	2.55	ug/l	94
20) Methylene Chloride	4.55	84	18358	1.09	ug/l	92
25) 2-Butanone	6.85	43	7850	1.45	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
Data File : VN049680.D
Acq On : 2 Jul 2018 19:55
Operator : MD\SY
Sample : PB110744 ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110744 ZHE#07

Quant Time: Jul 03 05:10:41 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

