

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
 Data File : VN049681.D
 Acq On : 2 Jul 2018 20:21
 Operator : MD\SY
 Sample : PB110744 ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110744 ZHE#08

Quant Time: Jul 03 05:12:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1251440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2072433	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1680278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	654649	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	869690	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
35) Dibromofluoromethane	7.59	113	719333	46.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.86%	
50) Toluene-d8	10.09	98	2740150	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
62) 4-Bromofluorobenzene	12.41	95	783713	37.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.70%	
Target Compounds						
16) Acetone	3.82	43	49049	7.01	ug/l	96
18) Methyl Acetate	4.33	43	115502	2.74	ug/l	100
20) Methylene Chloride	4.54	84	19452	1.14	ug/l	93
25) 2-Butanone	6.85	43	8286	1.51	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070218\
Data File : VN049681.D
Acq On : 2 Jul 2018 20:21
Operator : MD\SY
Sample : PB110744 ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110744 ZHE#08

Quant Time: Jul 03 05:12:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

