

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062344.D
 Acq On : 2 Jul 2020 19:42
 Operator : JC/MD
 Sample : PB129917ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#02

Quant Time: Jul 03 00:36:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	124153	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	255588	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	251978	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	85009	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	103923	59.65	ug/l	0.00
Spiked Amount			Recovery	=	119.30%	
35) Dibromofluoromethane	7.55	113	80501	49.09	ug/l	0.00
Spiked Amount			Recovery	=	98.18%	
50) Toluene-d8	10.06	98	335388	51.17	ug/l	0.00
Spiked Amount			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.38	95	106659	48.16	ug/l	0.00
Spiked Amount			Recovery	=	96.32%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.72	59	269	1.075	ug/l #	72
16) Acetone	3.78	43	9037	12.679	ug/l #	88
18) Methyl Acetate	4.28	43	2662	1.922	ug/l #	64
20) Methylene Chloride	4.50	84	6018	3.561	ug/l	91
25) 2-Butanone	6.80	43	1990	2.337	ug/l #	88
29) Tetrahydrofuran	7.17	42	563	1.009	ug/l #	31
31) Cyclohexane	7.62	56	3264	1.356	ug/l #	21
36) 1,1-Dichloropropene	7.63	75	11464	4.462	ug/l #	46
37) Ethyl Acetate	6.80	43	1990	0.935	ug/l #	66
38) Carbon Tetrachloride	7.63	117	13033	5.381	ug/l #	15
41) Methacrylonitrile	7.17	41	227	0.230	ug/l #	14
43) Isopropyl Acetate	8.13	43	47430	13.677	ug/l	98
45) 1,2-Dichloropropane	9.24	63	1324	0.685	ug/l #	43
48) Methyl methacrylate	9.24	41	88881	57.162	ug/l #	34
49) 1,4-Dioxane	9.24	88	70	2.030	ug/l #	1
51) 4-Methyl-2-Pentanone	9.95	43	4572	2.257	ug/l	98
54) cis-1,3-Dichloropropene	9.88	75	23055	7.443	ug/l #	67
56) Ethyl methacrylate	10.52	69	1080	0.431	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	4325	1.338	ug/l #	42
59) 2-Hexanone	10.74	43	58025	39.660	ug/l #	48
71) Bromoform	12.38	173	605	0.445	ug/l #	1
74) N-amyl acetate	12.01	43	2119	4.159	ug/l #	38
76) 1,2,3-Trichloropropane	12.37	75	55500	27.734	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.71	105	2232	0.418	ug/l	92
81) trans-1,4-Dichloro-2-buten	12.37	75	55500	80.993	ug/l #	13
95) Naphthalene	15.11	128	840	4.186	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062344.D
Acq On : 2 Jul 2020 19:42
Operator : JC/MD
Sample : PB129917ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02

Quant Time: Jul 03 00:36:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.00 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Instrument :

MSVOA_N

ClientSampleId :


PB129917ZHE#02

Tot Ion:168 Resp: 124153

Ion Ratio Lower Upper




168 100

99 63.5 45.8 68.8



Abundance Ion 167.90 (167.60 to 168.60): VN062059.D (-1800) (-)

Ion 98.90 (98.60 to 99.60): VN062059.D (-1800) (-)



#11

Tert butyl alcohol

Concen: 1.075 ug/l

RT: 4.72 min Scan# 916

Delta R.T. -0.02 min

Lab File: VN062344.D


Acq: 2 Jul 2020 19:42

Tgt Ion: 59 Resp: 269

Ion Ratio Lower Upper



59 100


57 0.0 8.3 12.5#



Abundance Ion 59.00 (58.70 to 59.70): VN062059.D (-895) (-)

Ion 57.00 (56.70 to 57.70): VN062059.D (-895) (-)





#16

Acetone

Concen: 12.679 ug/l

RT: 3.78 min Scan# 623

Delta R.T. 0.00 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Instrument :

MSVOA_N

ClientSampleId :





PB129917ZHE#02

Tot Ion: 43 Resp: 9037

Ion Ratio Lower Upper

43 100

58 41.6 27.5 41.3#



#18

Methyl Acetate

Concen: 1.922 ug/l

RT: 4.28 min Scan# 781

Delta R.T. 0.01 min

Lab File: VN062344.D




Acq: 2 Jul 2020 19:42


Tgt Ion: 43 Resp: 2662

Ion Ratio Lower Upper

43 100

74 9.6 22.9 34.3#









#20
Methvlene Chloride
Concen: 3.561 ug/l
RT: 4.50 min Scan# 849
Delta R.T. 0.00 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02

Tot Ion:	84	Resp:	6018
Ion	Ratio	Lower	Upper
84	100		
49	128.5	90.6	136.0
51	38.2	27.8	41.8
86	64.9	50.1	75.1




Abundance Ion 83.90 (83.60 to 84.60): VNO
Ion 48.90 (48.60 to 49.60): VNO
Ion 50.90 (50.60 to 51.60): VNO
Ion 85.90 (85.60 to 86.60): VNO






#25
2-Butanone
Concen: 2.337 ug/l
RT: 6.80 min Scan# 1563
Delta R.T. 0.02 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Tgt Ion:	43	Resp:	1990
Ion	Ratio	Lower	Upper
43	100		
72	37.4	24.8	37.2#



Abundance Ion 43.00 (42.70 to 43.70): VNO
Ion 72.00 (71.70 to 72.70): VNO









#29
Tetrahydrofuran
Concen: 1.009 ug/l
RT: 7.17 min Scan# 1680
Delta R.T. 0.01 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42




Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02


Tot Ion: 42 Resp: 563
Ion Ratio Lower Upper
42 100
72 0.0 41.1 61.7#
71 5.3 38.2 57.4#



#31
Cyclohexane
Concen: 1.356 ug/l
RT: 7.62 min Scan# 1818
Delta R.T. 0.01 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Tgt Ion: 56 Resp: 3264
Ion Ratio Lower Upper
56 100
69 144.5 28.0 42.0#
84 132.6 76.1 114.1#









#33
 1,2-Dichloroethane-d4
 Concen: 59.648 ug/l
 RT: 7.99 min Scan# 1932
 Delta R.T. -0.00 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#02


Tot Ion: 65 Resp: 103923
 Ion Ratio Lower Upper
 65 100
 67 52.7 0.0 106.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2108
 Delta R.T. -0.00 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42

Tgt Ion: 114 Resp: 255588
 Ion Ratio Lower Upper
 114 100
 63 21.1 0.0 39.2
 88 15.7 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 49.094 ug/l
 RT: 7.55 min Scan# 1796
 Delta R.T. -0.00 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#02


Tot Ion	Ratio	Lower	Upper
113	100		
111	103.5	82.2	123.4
192	16.8	14.0	21.0



#36
 1,1-Dichloropropene
 Concen: 4.462 ug/l
 RT: 7.63 min Scan# 1820
 Delta R.T. -0.13 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42

Tgt Ion	Ratio	Lower	Upper
75	100		
110	4.9	17.9	53.8#
77	0.0	24.9	37.3#









#37
Ethyl Acetate
Concen: 0.935 ug/l
RT: 6.80 min Scan# 1563
Delta R.T. -0.08 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42




Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02


Tot Ion: 43 Resp: 1990
Ion Ratio Lower Upper
43 100
61 0.0 12.0 18.0#
70 0.0 9.6 14.4#



#38
Carbon Tetrachloride
Concen: 5.381 ug/l
RT: 7.63 min Scan# 1820
Delta R.T. -0.11 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Tgt Ion: 117 Resp: 13033
Ion Ratio Lower Upper
117 100
119 4.7 78.2 117.2#
121 0.0 25.8 38.6#









#41
Methacrylonitrile
Concen: 0.230 ug/l
RT: 7.17 min Scan# 1677
Delta R.T. 0.04 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42




Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02


Tot Ion:	41	Resp:	227
Ion	Ratio	Lower	Upper
41	100		
39	0.0	44.2	66.2#
67	0.0	79.5	119.3#
52	0.0	32.3	48.5#



#43
Isopropyl Acetate
Concen: 13.677 ug/l
RT: 8.13 min Scan# 1976
Delta R.T. -0.00 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Tgt Ion:	43	Resp:	47430
Ion	Ratio	Lower	Upper
43	100		
61	21.1	17.2	25.8
87	13.8	12.1	18.1





#45

1,2-Dichloropropane

Concen: 0.685 ug/l

RT: 9.24 min Scan# 2321

Delta R.T. 0.15 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Instrument :

MSVOA_N

ClientSampleId :


PB129917ZHE#02

Tot Ion: 63 Resp: 1324

Ion Ratio Lower Upper

63 100

65 0.0 25.0 37.4#



Abundance Ion 62.90 (62.60 to 63.60): VNO

Ion 64.90 (64.60 to 65.60): VNO

9.24

800

600



400

200

0

9.20 9.25

Time-->



#48

Methyl methacrylate

Concen: 57.162 ug/l

RT: 9.24 min Scan# 2321

Delta R.T. 0.07 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42


Tgt Ion: 41 Resp: 88881

Ion Ratio Lower Upper

41 100

69 0.0 80.7 121.1#

39 55.5 54.5 81.7



Abundance Ion 41.00 (40.70 to 41.70): VNO

Ion 69.00 (68.70 to 69.70): VNO

Ion 39.00 (38.70 to 39.70): VNO

60000

50000

40000

30000


20000


10000

0

9.15 9.20 9.25 9.30

Time-->





#49

1,4-Dioxane

Concen: 2.030 ug/l

RT: 9.24 min Scan# 2322

Delta R.T. 0.07 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Instrument :

MSVOA_N

ClientSampleId :

PB129917ZHE#02


Tot Ion: 88 Resp: 70

Ion Ratio Lower Upper

88 100

43 1261150.0 22.0 33.0#




58 6872.9 49.2 73.8#



Abundance Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 51.169 ug/l

RT: 10.06 min Scan# 2577

Delta R.T. -0.00 min

Lab File: VN062344.D


Acq: 2 Jul 2020 19:42

Tgt Ion: 98 Resp: 335388

Ion Ratio Lower Upper



98 100


100 65.0 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0









#51
 4-Methyl-2-Pentanone
 Concen: 2.257 ug/l
 RT: 9.95 min Scan# 2544
 Delta R.T. -0.00 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#02


Tot Ion: 43 Resp: 4572
 Ion Ratio Lower Upper
 43 100
 58 44.0 34.2 51.2



#54
 cis-1,3-Dichloropropene
 Concen: 7.443 ug/l
 RT: 9.88 min Scan# 2520
 Delta R.T. 0.07 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42

Tgt Ion: 75 Resp: 23055
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.0 37.6#
 39 35.0 36.9 55.3#





#56

Ethyl methacrylate

Concen: 0.431 ug/l

RT: 10.52 min Scan# 2720

Delta R.T. 0.12 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Instrument :

MSVOA_N

ClientSampleId :

PB129917ZHE#02


Tot Ion: 69 Resp: 1080

Ion Ratio Lower Upper

69 100

41 4800.8 46.6 70.0#

39 2209.9 31.0 46.4#






Abundance

Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0



#57

1,3-Dichloropropane

Concen: 1.338 ug/l

RT: 10.74 min Scan# 2788

Delta R.T. 0.06 min

Lab File: VN062344.D


Acq: 2 Jul 2020 19:42

Tgt Ion: 76 Resp: 4325

Ion Ratio Lower Upper

76 100



78 0.0 25.8 38.8#




Abundance

Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0









#59
2-Hexanone
Concen: 39.660 ug/l
RT: 10.74 min Scan# 2788
Delta R.T. 0.02 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42




Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02


Tot Ion: 43 Resp: 58025
Ion Ratio Lower Upper
43 100
58 19.6 29.5 88.5#



#62
4-Bromofluorobenzene
Concen: 48.164 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. -0.00 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Tgt Ion: 95 Resp: 106659
Ion Ratio Lower Upper
95 100
174 71.8 0.0 154.8
176 68.7 0.0 147.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. -0.00 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02

Tot Ion:117 Resp: 251978
Ion Ratio Lower Upper
117 100
82 58.1 43.8 65.8
119 32.5 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 0.445 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.28 min
Lab File: VN062344.D
Acq: 2 Jul 2020 19:42

Tgt Ion:173 Resp: 605
Ion Ratio Lower Upper
173 100
175 800.5 24.7 74.1#
254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. 0.00 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Instrument :

MSVOA_N

ClientSampleId :

PB129917ZHE#02


Tot Ion:152 Resp: 85009

Ion Ratio Lower Upper

152 100

115 60.6 29.5 88.5




150 158.6 0.0 351.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 4.159 ug/l

RT: 12.01 min Scan# 3184

Delta R.T. -0.04 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Tgt Ion: 43 Resp: 2119


Ion Ratio Lower Upper

43 100

70 0.0 39.4 59.2#

55 0.0 21.7 32.5#

61 0.0 21.3 31.9#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V









#76
 1,2,3-Trichloropropane
 Concen: 27.734 ug/l
 RT: 12.37 min Scan# 3297
 Delta R.T. -0.15 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#02


Tot Ion: 75 Resp: 55500
 Ion Ratio Lower Upper
 75 100
 77 1.2 0.0 0.0#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.418 ug/l
 RT: 12.71 min Scan# 3401
 Delta R.T. -0.00 min
 Lab File: VN062344.D
 Acq: 2 Jul 2020 19:42

Tgt Ion:105 Resp: 2232
 Ion Ratio Lower Upper
 105 100
 120 44.0 24.6 74.0





#81

trans-1,4-Dichloro-2-butene

Concen: 80.993 ug/l

RT: 12.37 min Scan# 3297

Delta R.T. 0.10 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42

Instrument :

MSVOA_N

ClientSampleId :

PB129917ZHE#02


Tot Ion: 75 Resp: 55500

Ion Ratio Lower Upper

75 100

53 0.0 72.0 108.0#




89 0.0 36.9 55.3#



Abundance Ion 74.90 (74.60 to 75.60): VN062059.D (-3256) (-)

Ion 53.00 (52.70 to 53.70): VN062059.D (-3256) (-)

Ion 88.90 (88.60 to 89.60): VN062059.D (-3256) (-)



#95

Naphthalene

Concen: 4.186 ug/l

RT: 15.11 min Scan# 4148

Delta R.T. 0.01 min

Lab File: VN062344.D

Acq: 2 Jul 2020 19:42


Tgt Ion: 128 Resp: 840

Ion Ratio Lower Upper

128 100

127 0.0 10.5 15.7#

129 0.0 8.6 13.0#



Abundance Ion 128.00 (127.70 to 128.70): VN062059.D (-4131) (-)

Ion 127.00 (126.70 to 127.70): VN062059.D (-4131) (-)

Ion 129.00 (128.70 to 129.70): VN062059.D (-4131) (-)

