

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062343.D
 Acq On : 2 Jul 2020 19:18
 Operator : JC/MD
 Sample : PB129917ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#01

Manual Integrations
 APPROVED

sam
 7/6/2020 11:34:16 AM

Quant Time: Jul 03 06:09:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	122499	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	245016	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	242390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	81140	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	102043	59.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.72%	
35) Dibromofluoromethane	7.55	113	78972	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
50) Toluene-d8	10.06	98	331157	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	12.38	95	102301	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
Target Compounds						
16) Acetone	3.78	43	8996	12.829	ug/l	91
18) Methyl Acetate	4.28	43	2526m	1.848	ug/l	
20) Methylene Chloride	4.50	84	6193	3.724	ug/l	93
43) Isopropyl Acetate	8.13	43	49165	14.789	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062343.D
Acq On : 2 Jul 2020 19:18
Operator : JC/MD
Sample : PB129917ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#01

Manual Integrations
APPROVED

sam
7/6/2020 11:34:16 AM

Quant Time: Jul 03 06:09:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

