

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062344.D
 Acq On : 2 Jul 2020 19:42
 Operator : JC/MD
 Sample : PB129917ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#02

Quant Time: Jul 03 03:46:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	124153	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	255588	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	251978	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	85009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	103923	59.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.30%	
35) Dibromofluoromethane	7.55	113	80501	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	10.06	98	335388	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	12.38	95	106659	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
Target Compounds						
16) Acetone	3.78	43	9037	12.679	ug/l #	88
18) Methyl Acetate	4.28	43	2662	1.922	ug/l #	64
20) Methylene Chloride	4.50	84	6018	3.561	ug/l	91
43) Isopropyl Acetate	8.13	43	47430	13.677	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062344.D
Acq On : 2 Jul 2020 19:42
Operator : JC/MD
Sample : PB129917ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#02

Quant Time: Jul 03 03:46:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

