

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062346.D
 Acq On : 2 Jul 2020 20:31
 Operator : JC/MD
 Sample : PB129917ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#04

Quant Time: Jul 03 03:54:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	123989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	261952	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	257264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	90932	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	103563	59.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.04%	
35) Dibromofluoromethane	7.55	113	82210	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	10.06	98	344073	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
62) 4-Bromofluorobenzene	12.38	95	111235	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
Target Compounds						
16) Acetone	3.78	43	8932	12.506	ug/l	93
18) Methyl Acetate	4.28	43	3841	2.777	ug/l #	80
20) Methylene Chloride	4.51	84	6006	3.559	ug/l	93
43) Isopropyl Acetate	8.13	43	47002	13.224	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062346.D
Acq On : 2 Jul 2020 20:31
Operator : JC/MD
Sample : PB129917ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#04

Quant Time: Jul 03 03:54:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

