

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062347.D
 Acq On : 2 Jul 2020 20:55
 Operator : JC/MD
 Sample : PB129917ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#05

Quant Time: Jul 03 03:56:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	124663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	260584	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	253016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	86560	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105506	60.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.62%	
35) Dibromofluoromethane	7.55	113	81794	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
50) Toluene-d8	10.06	98	342217	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.38	95	106844	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
Target Compounds						
16) Acetone	3.79	43	8628	11.853	ug/l	93
18) Methyl Acetate	4.29	43	4408	3.170	ug/l #	74
20) Methylene Chloride	4.51	84	6425	3.800	ug/l	97
43) Isopropyl Acetate	8.12	43	46791	13.234	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062347.D
Acq On : 2 Jul 2020 20:55
Operator : JC/MD
Sample : PB129917ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#05

Quant Time: Jul 03 03:56:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

