

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062348.D
 Acq On : 2 Jul 2020 21:19
 Operator : JC/MD
 Sample : PB129917ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#06

Quant Time: Jul 03 03:58:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	125543	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	268181	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	265151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92829	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	106430	60.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.82%	
35) Dibromofluoromethane	7.55	113	83895	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
50) Toluene-d8	10.06	98	352213	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
62) 4-Bromofluorobenzene	12.38	95	112484	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
Target Compounds						
16) Acetone	3.77	43	9012	12.447	ug/l	89
18) Methyl Acetate	4.28	43	4339	3.098	ug/l #	84
20) Methylene Chloride	4.50	84	6394	3.753	ug/l	95
43) Isopropyl Acetate	8.13	43	46415	12.756	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062348.D
Acq On : 2 Jul 2020 21:19
Operator : JC/MD
Sample : PB129917ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#06

Quant Time: Jul 03 03:58:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

