

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062349.D
 Acq On : 2 Jul 2020 21:43
 Operator : JC/MD
 Sample : PB129917ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129917ZHE#07

Quant Time: Jul 03 04:00:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	115459	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	244972	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	237812	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	77495	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	98865	61.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.04%	
35) Dibromofluoromethane	7.55	113	77362	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	10.06	98	321884	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
62) 4-Bromofluorobenzene	12.38	95	98820	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
Target Compounds						
16) Acetone	3.79	43	6727	9.327	ug/l	94
18) Methyl Acetate	4.28	43	4190	3.253	ug/l #	89
20) Methylene Chloride	4.49	84	5576	3.547	ug/l	84
43) Isopropyl Acetate	8.13	43	41530	12.495	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062349.D
Acq On : 2 Jul 2020 21:43
Operator : JC/MD
Sample : PB129917ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129917ZHE#07

Quant Time: Jul 03 04:00:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

