

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070220\
 Data File : VN062353.D
 Acq On : 2 Jul 2020 23:20
 Operator : JC/MD
 Sample : L3188-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-070120

Quant Time: Jul 03 04:20:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	133059	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	275238	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	276451	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97220	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	109955	58.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.78%	
35) Dibromofluoromethane	7.55	113	86973	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
50) Toluene-d8	10.06	98	368421	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	12.38	95	114988	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
Target Compounds						
16) Acetone	3.78	43	9808	12.892	ug/l	98
18) Methyl Acetate	4.28	43	4334	2.920	ug/l #	64
20) Methylene Chloride	4.51	84	8655	4.853	ug/l	95
43) Isopropyl Acetate	8.13	43	39662	10.621	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070220\
Data File : VN062353.D
Acq On : 2 Jul 2020 23:20
Operator : JC/MD
Sample : L3188-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-070120

Quant Time: Jul 03 04:20:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

