

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
 Data File : VN067617.D
 Acq On : 2 Jul 2021 15:28
 Operator : JC/MD
 Sample : M2913-05ME
 Misc : 6.32g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 D-3100ME

Quant Time: Jul 05 05:00:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 12:55:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.078	168	353404	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.965	114	647697	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	613014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	252878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	264722	50.058	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.120%
35) Dibromofluoromethane	8.019	113	189564	47.319	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.640%
50) Toluene-d8	10.443	98	819773	48.485	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.960%
62) 4-Bromofluorobenzene	12.734	95	285737	47.372	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.740%
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.980	164	125463	28.753	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
Data File : VN067617.D
Acq On : 2 Jul 2021 15:28
Operator : JC/MD
Sample : M2913-05ME
Misc : 6.32g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
D-3100ME

Quant Time: Jul 05 05:00:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 12:55:54 2021
Response via : Initial Calibration

