

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
 Data File : VN067620.D
 Acq On : 2 Jul 2021 16:40
 Operator : JC/MD
 Sample : M2852-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MANHOLE-WC-D

Quant Time: Jul 05 05:02:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 12:55:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	382631	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	695149	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	695150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	255040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	277541	48.473	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.940%
35) Dibromofluoromethane	8.024	113	209568	48.742	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.480%
50) Toluene-d8	10.443	98	901637	49.687	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.380%
62) 4-Bromofluorobenzene	12.736	95	308386	47.637	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.280%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	22370	12.324	ug/l	90
20) Methylene Chloride	5.092	84	123036	27.188	ug/l	95
43) Isopropyl Acetate	8.566	43	47504	4.329	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
Data File : VN067620.D
Acq On : 2 Jul 2021 16:40
Operator : JC/MD
Sample : M2852-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MANHOLE-WC-D

Quant Time: Jul 05 05:02:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 12:55:54 2021
Response via : Initial Calibration

