

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
 Data File : VN067624.D
 Acq On : 2 Jul 2021 18:18
 Operator : JC/MD
 Sample : M2896-04RE 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FS-02RE

Manual Integrations
 APPROVED

MMDadoda
 7/7/2021 8:51:41 AM

Quant Time: Jul 05 05:03:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 12:55:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	379788	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	720197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	783857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	420003	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	278800	49.058	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.120%		
35) Dibromofluoromethane	8.024	113	205589	46.153	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.300%		
50) Toluene-d8	10.443	98	955745	50.836	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.680%		
62) 4-Bromofluorobenzene	12.733	95	439424	65.519	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	131.040%		
Target Compounds						Qvalue
16) Acetone	4.291	43	227306	126.160	ug/l	93
18) Methyl Acetate	4.856	43	575650	125.405	ug/l	96
20) Methylene Chloride	5.098	84	5399m	1.202	ug/l	
25) 2-Butanone	7.340	43	914889	314.703	ug/l	95
29) Tetrahydrofuran	7.689	42	198643	102.585	ug/l	94
40) Benzene	8.461	78	464308	23.039	ug/l	97
51) 4-Methyl-2-Pentanone	10.333	43	154781	22.268	ug/l	98
52) Toluene	10.507	92	1484436	110.596	ug/l	99
59) 2-Hexanone	11.089	43	939180	192.764	ug/l	98
67) Ethyl Benzene	11.848	91	544187	18.672	ug/l	97
68) m/p-Xylenes	11.953	106	504498	45.172	ug/l	89
69) o-Xylene	12.283	106	329875	29.828	ug/l	88
78) n-propylbenzene	12.924	91	60773	1.566	ug/l	92
80) 1,3,5-Trimethylbenzene	13.063	105	65459	2.323	ug/l	94
84) 1,2,4-Trimethylbenzene	13.372	105	248399	9.008	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
Data File : VN067624.D
Acq On : 2 Jul 2021 18:18
Operator : JC/MD
Sample : M2896-04RE 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FS-02RE

Manual Integrations
APPROVED

MMDadoda
7/7/2021 8:51:41 AM

Quant Time: Jul 05 05:03:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 12:55:54 2021
Response via : Initial Calibration

