

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
 Data File : VN067626.D
 Acq On : 2 Jul 2021 19:07
 Operator : JC/MD
 Sample : M2908-02 100X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-CROPSEY-AQUEOUS

Quant Time: Jul 05 05:04:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 12:55:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	375598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	695700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	648404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	254903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	275016	48.932	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.860%
35) Dibromofluoromethane	8.024	113	204426	47.508	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.020%
50) Toluene-d8	10.443	98	872981	48.069	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.140%
62) 4-Bromofluorobenzene	12.733	95	290307	44.809	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.620%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	9076	5.094	ug/l	100
18) Methyl Acetate	4.865	43	8578	1.890	ug/l #	56
95) Naphthalene	15.509	128	9033	5.253	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
Data File : VN067626.D
Acq On : 2 Jul 2021 19:07
Operator : JC/MD
Sample : M2908-02 100X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-CROPSEY-AQUEOUS

Quant Time: Jul 05 05:04:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 12:55:54 2021
Response via : Initial Calibration

