

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
 Data File : VN067627.D
 Acq On : 2 Jul 2021 19:31
 Operator : JC/MD
 Sample : M2896-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FS-03

Manual Integrations
 APPROVED

MMDadoda
 7/7/2021 8:51:48 AM

Quant Time: Jul 05 05:05:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 12:55:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	369499	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	688344	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	704332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	278246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	269630	48.765	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.540%		
35) Dibromofluoromethane	8.024	113	202350	47.528	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.060%		
50) Toluene-d8	10.446	98	898855	50.023	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.040%		
62) 4-Bromofluorobenzene	12.733	95	312207	48.704	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.400%		
Target Compounds					Qvalue	
16) Acetone	4.291	43	85662	48.868	ug/l	91
18) Methyl Acetate	4.878	43	9313m	2.085	ug/l	
20) Methylene Chloride	5.101	84	28591	6.543	ug/l #	84
25) 2-Butanone	7.351	43	18397	6.504	ug/l	100
43) Isopropyl Acetate	8.566	43	41154	3.788	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
Data File : VN067627.D
Acq On : 2 Jul 2021 19:31
Operator : JC/MD
Sample : M2896-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FS-03

Manual Integrations
APPROVED

MMDadoda
7/7/2021 8:51:48 AM

Quant Time: Jul 05 05:05:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 12:55:54 2021
Response via : Initial Calibration

