

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
 Data File : VN067628.D
 Acq On : 2 Jul 2021 19:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Manual Integrations
 APPROVED

MMDadoda
 7/7/2021 8:51:52 AM

Quant Time: Jul 05 05:05:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 12:55:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	351949	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	650747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	607091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	237626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	255972	48.603	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.200%		
35) Dibromofluoromethane	8.024	113	189291	47.030	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.060%		
50) Toluene-d8	10.443	98	820304	48.289	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.580%		
62) 4-Bromofluorobenzene	12.733	95	273206	45.083	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.160%		
Target Compounds						
18) Methyl Acetate	4.865	43	9348	2.198	ug/l	# 63
95) Naphthalene	15.504	128	4292m	4.932	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
Data File : VN067628.D
Acq On : 2 Jul 2021 19:56
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Manual Integrations
APPROVED

MMDadoda
7/7/2021 8:51:52 AM

Quant Time: Jul 05 05:05:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 12:55:54 2021
Response via : Initial Calibration

