



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070221\
 Data File : VN067603.D
 Acq On : 2 Jul 2021 8:46
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070221W.M
 Title : SW846 8260
 Last Update : Fri Jul 02 12:55:54 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.7	20232	PASS
75	95	30	60	59.0	50333	PASS
95	95	100	100	100.0	85267	PASS
96	95	5	9	6.8	5765	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.6	63584	PASS
175	174	5	9	8.1	5178	PASS
176	174	95	101	95.5	60712	PASS
177	176	5	9	6.8	4112	PASS