

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
 Data File : VN049699.D
 Acq On : 3 Jul 2018 17:39
 Operator : MD\SY
 Sample : PB110796ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110796ZHE#02

Quant Time: Jul 04 00:21:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1283129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2093224	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1748883	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	695670	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	888201	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
35) Dibromofluoromethane	7.59	113	734777	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
50) Toluene-d8	10.09	98	2803409	45.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.54%	
62) 4-Bromofluorobenzene	12.41	95	837194	40.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.06%	
Target Compounds						
16) Acetone	3.82	43	46540	5.874	ug/l	93
18) Methyl Acetate	4.33	43	106744	1.869	ug/l	98
20) Methylene Chloride	4.55	84	28110	1.611	ug/l	95
25) 2-Butanone	6.85	43	24255	4.312	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070318\
Data File : VN049699.D
Acq On : 3 Jul 2018 17:39
Operator : MD\SY
Sample : PB110796ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110796ZHE#02

Quant Time: Jul 04 00:21:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

