

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
 Data File : VN049700.D
 Acq On : 3 Jul 2018 18:05
 Operator : MD\SY
 Sample : PB110796ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110796ZHE#03

Quant Time: Jul 04 00:26:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1254846	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2024057	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1672106	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	664707	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	862755	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
35) Dibromofluoromethane	7.59	113	717403	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
50) Toluene-d8	10.09	98	2694138	45.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.98%	
62) 4-Bromofluorobenzene	12.40	95	801795	39.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.30%	
Target Compounds						
16) Acetone	3.82	43	43891	5.370	ug/l	95
18) Methyl Acetate	4.33	43	101224	1.627	ug/l	97
20) Methylene Chloride	4.55	84	29448	1.726	ug/l	87
25) 2-Butanone	6.85	43	23944	4.352	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
Data File : VN049700.D
Acq On : 3 Jul 2018 18:05
Operator : MD\SY
Sample : PB110796ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110796ZHE#03

Quant Time: Jul 04 00:26:39 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

