

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
 Data File : VN049704.D
 Acq On : 3 Jul 2018 19:49
 Operator : MD\SY
 Sample : PB110796ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110796ZHE#07

Quant Time: Jul 04 00:35:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1177855	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1940864	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1572968	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	607939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	815085	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
35) Dibromofluoromethane	7.59	113	678502	46.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.52%	
50) Toluene-d8	10.09	98	2565307	45.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.34%	
62) 4-Bromofluorobenzene	12.40	95	742416	38.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.58%	
Target Compounds						
16) Acetone	3.83	43	41486	5.465	ug/l	96
20) Methylene Chloride	4.55	84	26949	1.682	ug/l	98
25) 2-Butanone	6.85	43	19957	3.865	ug/l	96
59) 2-Hexanone	10.76	43	8783	1.064	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070318\
Data File : VN049704.D
Acq On : 3 Jul 2018 19:49
Operator : MD\SY
Sample : PB110796ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110796ZHE#07

Quant Time: Jul 04 00:35:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

