

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
 Data File : VN049705.D
 Acq On : 3 Jul 2018 20:15
 Operator : MD\SY
 Sample : PB110796ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110796ZHE#08

Quant Time: Jul 04 00:38:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1238353	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2016330	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1643085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	607477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	857561	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
35) Dibromofluoromethane	7.59	113	713202	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
50) Toluene-d8	10.09	98	2672538	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
62) 4-Bromofluorobenzene	12.40	95	758674	37.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.32%	
Target Compounds						
16) Acetone	3.82	43	41967	4.946	ug/l	94
18) Methyl Acetate	4.33	43	103309	1.891	ug/l	99
20) Methylene Chloride	4.55	84	26891	1.597	ug/l #	89
25) 2-Butanone	6.85	43	21180	3.901	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
Data File : VN049705.D
Acq On : 3 Jul 2018 20:15
Operator : MD\SY
Sample : PB110796ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110796ZHE#08

Quant Time: Jul 04 00:38:30 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

