

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
 Data File : VN049707.D
 Acq On : 3 Jul 2018 21:06
 Operator : MD\SY
 Sample : PB110796ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110796ZHE#10

Quant Time: Jul 04 00:44:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 30 00:55:30 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1190604	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1986002	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1649209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	612967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	839420	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
35) Dibromofluoromethane	7.59	113	698244	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
50) Toluene-d8	10.09	98	2635017	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
62) 4-Bromofluorobenzene	12.41	95	771513	38.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.76%	
Target Compounds						
16) Acetone	3.82	43	42848	5.764	ug/l	97
18) Methyl Acetate	4.33	43	96984	1.703	ug/l	95
20) Methylene Chloride	4.55	84	28790	1.778	ug/l	89
25) 2-Butanone	6.85	43	20933	4.010	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070318\
Data File : VN049707.D
Acq On : 3 Jul 2018 21:06
Operator : MD\SY
Sample : PB110796ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110796ZHE#10

Quant Time: Jul 04 00:44:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 30 00:55:30 2018
Response via : Initial Calibration

