


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN070318\
 Data File : VN049684.D
 Acq On : 3 Jul 2018 9:10
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062918W.M
 Title : SW846 8260
 Last Update : Sat Jun 30 00:55:30 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3348

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	23893	PASS
75	95	30	60	48.5	61293	PASS
95	95	100	100	100.0	126362	PASS
96	95	5	9	6.8	8623	PASS
173	174	0.00	2	1.8	1929	PASS
174	95	50	100	85.4	107893	PASS
175	174	5	9	7.9	8504	PASS
176	174	95	101	98.4	106122	PASS
177	176	5	9	6.5	6915	PASS