

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN070319\
 Data File : VN056580.D
 Acq On : 3 Jul 2019 11:06
 Operator : JC/SP
 Sample : VSTDICV020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN070319

Quant Time: Jul 03 11:29:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N070319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 03 11:05:14 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	101	0.00
2 M	Dichlorodifluoromethane	1.504	1.528	-1.6	103	0.00
3 M	Chloromethane	2.183	2.100	3.8	100	0.00
4 M	Vinyl Chloride	2.055	2.099	-2.1	107	0.00
5 M	Bromomethane	1.096	1.232	-12.4	109	0.00
6 M	Chloroethane	1.138	1.145	-0.6	105	0.00
7 M	Trichlorofluoromethane	2.638	2.667	-1.1	104	0.00
8 T	Diethyl Ether	1.119	1.127	-0.7	105	0.00
9	1,1,2-Trichlorotrifluoroeth	1.671	1.651	1.2	103	0.00
10 M	1,1-Dichloroethene	1.640	1.638	0.1	103	0.00
11	Methyl Iodide	2.345	2.566	-9.4	121	0.00
12	Methyl Acetate	2.212	2.152	2.7	107	0.00
13 M	Acrolein	0.216	0.216	0.0	105	0.00
14 M	Acrylonitrile	0.971	0.967	0.4	104	0.00
15 M	Acetone	0.271	0.273	-0.7	102	0.00
16 M	Carbon Disulfide	3.943	4.075	-3.3	114	0.00
17	Allyl chloride	2.870	2.773	3.4	104	0.00
18 M	Methylene Chloride	1.939	1.927	0.6	104	0.00
19 M	trans-1,2-Dichloroethene	1.779	1.773	0.3	104	0.00
20 T	Diisopropyl ether	5.929	5.975	-0.8	104	0.00
21 M	1,1-Dichloroethane	3.345	3.311	1.0	102	0.00
22 M	cis-1,2-Dichloroethene	2.092	2.076	0.8	104	0.00
23 M	tert-Butyl Alcohol	0.363	0.372	-2.5	117	0.00
24 M	Methyl tert-Butyl Ether	5.424	5.447	-0.4	106	0.00
25 M	Chloroform	3.278	3.240	1.2	102	0.00
26	Cyclohexane	3.081	3.037	1.4	103	0.00
27 s	1,2-Dichloroethane-d4	2.068	2.102	-1.6	102	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
29	1,1-Dichloropropene	0.443	0.433	2.3	106	0.00
30 M	2-Butanone	0.246	0.238	3.3	103	0.00
31	2,2-Dichloropropane	0.426	0.400	6.1	104	0.00
32 M	1,1,1-Trichloroethane	0.472	0.447	5.3	103	0.00
33 M	Carbon Tetrachloride	0.400	0.373	6.8	103	0.00
34 M	Benzene	1.385	1.327	4.2	104	0.00
35	Methacrylonitrile	0.218	0.219	-0.5	103	0.00
36 M	1,2-Dichloroethane	0.450	0.436	3.1	105	0.00
37 M	Trichloroethene	0.361	0.346	4.2	105	0.00
38	Methylcyclohexane	0.555	0.542	2.3	107	0.00
39 M	1,2-Dichloropropane	0.373	0.358	4.0	103	0.00
40	Dibromomethane	0.231	0.221	4.3	104	0.00
41 M	Bromodichloromethane	0.427	0.392	8.2	102	0.00
42 M	Vinyl Acetate	0.874	0.850	2.7	106	0.00
43	Ethyl Acetate	0.475	0.455	4.2	103	0.00
44	Isopropyl Acetate	0.741	0.696	6.1	103	0.00
45 T	1,4-Dioxane	0.007	0.007#	0.0	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN070319\
 Data File : VN056580.D
 Acq On : 3 Jul 2019 11:06
 Operator : JC/SP
 Sample : VSTDICV020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN070319

Quant Time: Jul 03 11:29:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N070319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 03 11:05:14 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.364	0.341	6.3	105	0.00
47	n-amyl Acetate	0.616	0.579	6.0	107	0.00
48 M	t-1,3-Dichloropropene	0.458	0.432	5.7	108	0.00
49 T	cis-1,3-Dichloropropene	0.528	0.500	5.3	107	0.00
50 M	1,1,2-Trichloroethane	0.340	0.326	4.1	104	0.00
51	Ethyl methacrylate	0.535	0.515	3.7	105	0.00
52	1,3-Dichloropropane	0.577	0.550	4.7	104	0.00
53 M	Dibromochloromethane	0.333	0.308	7.5	104	0.00
54 M	1,2-Dibromoethane	0.354	0.339	4.2	106	0.00
55 M	2-Chloroethyl vinyl ether	0.273	0.269	1.5	108	0.00
56 M	Bromoform	0.242	0.216	10.7	106	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
58 M	4-Methyl-2-Pentanone	0.539	0.534	0.9	106	0.00
59 M	2-Hexanone	0.400	0.393	1.8	106	0.00
60 S	4-Bromofluorobenzene	0.461	0.459	0.4	107	0.00
61 M	Tetrachloroethene	0.377	0.386	-2.4	108	0.00
62 M	Toluene	1.674	1.634	2.4	104	0.00
63 S	Toluene-d8	1.394	1.400	-0.4	103	0.00
64 M	Chlorobenzene	1.009	0.995	1.4	107	0.00
65	1,1,1,2-Tetrachloroethane	0.363	0.355	2.2	108	0.00
66 M	Ethyl Benzene	1.815	1.759	3.1	105	0.00
67 M	m/p-Xylenes	0.670	0.660	1.5	107	0.00
68 M	o-Xylene	0.663	0.639	3.6	104	0.00
69 M	Styrene	1.102	1.063	3.5	105	0.00
70	Isopropylbenzene	1.729	1.712	1.0	106	0.00
71 M	1,1,2,2-Tetrachloroethane	0.572	0.556	2.8	105	0.00
72	1,2,3-Trichloropropane	0.481	0.454	5.6	97	0.00
73	Bromobenzene	0.446	0.435	2.5	106	0.00
74	n-propylbenzene	1.961	1.923	1.9	108	0.00
75	2-Chlorotoluene	1.173	1.156	1.4	107	0.00
76	1,3,5-Trimethylbenzene	1.464	1.441	1.6	107	0.00
77	t-1,4-Dichloro-2-butene	0.154	0.137	11.0	109	0.00
78	4-Chlorotoluene	1.141	1.126	1.3	108	0.00
79	tert-butylbenzene	1.251	1.222	2.3	107	0.00
80	1,2,4-Trimethylbenzene	1.446	1.426	1.4	108	0.00
81	sec-Butylbenzene	1.659	1.641	1.1	108	0.00
82	p-Isopropyltoluene	1.491	1.470	1.4	108	0.00
83 M	1,3-Dichlorobenzene	0.740	0.731	1.2	111	0.00
84 M	1,4-Dichlorobenzene	0.714	0.721	-1.0	114	0.00
85	n-Butylbenzene	1.253	1.214	3.1	110	0.00
86 T	Hexachloroethane	0.225	0.212	5.8	110	0.00
87 M	1,2-Dichlorobenzene	0.752	0.752	0.0	109	0.00
88	1,2-Dibromo-3-Chloropropane	0.102	0.097	4.9	111	0.00
89	1,2,4-Trichlorobenzene	0.416	0.441	-6.0	132	0.00
90	Hexachlorobutadiene	0.258	0.264	-2.3	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN070319\
Data File : VN056580.D
Acq On : 3 Jul 2019 11:06
Operator : JC/SP
Sample : VSTDICV020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN070319

Quant Time: Jul 03 11:29:13 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N070319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 03 11:05:14 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.166	1.312	-12.5	140	0.00
92	1,2,3-Trichlorobenzene	0.436	0.494	-13.3	136	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0