

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070324\
 Data File : VN082794.D
 Acq On : 03 Jul 2024 11:27
 Operator : JC\MD
 Sample : VN0703WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0703WBL01

Quant Time: Jul 03 23:54:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	118848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	220282	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	211869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92928	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	89666	48.923	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.840%
35) Dibromofluoromethane	8.171	113	72624	50.074	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.140%
50) Toluene-d8	10.565	98	275242	50.746	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.500%
62) 4-Bromofluorobenzene	12.847	95	99595	53.469	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	106.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070324\
Data File : VN082794.D
Acq On : 03 Jul 2024 11:27
Operator : JC\MD
Sample : VN0703WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0703WBL01

Quant Time: Jul 03 23:54:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

