

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070324\
 Data File : VN082799.D
 Acq On : 03 Jul 2024 13:38
 Operator : JC\MD
 Sample : PB161888ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161888ZHE#02

Quant Time: Jul 03 23:58:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	117392	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	220829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	221328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	90062	49.748	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.500%		
35) Dibromofluoromethane	8.171	113	74788	51.439	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.880%		
50) Toluene-d8	10.571	98	282233	51.906	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.820%		
62) 4-Bromofluorobenzene	12.847	95	102260	54.764	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery =	109.520%		
Target Compounds						Qvalue
16) Acetone	4.436	43	36987	53.925	ug/l	98
20) Methylene Chloride	5.283	84	7166	4.129	ug/l	87
43) Isopropyl Acetate	8.694	43	155749	39.658	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070324\
Data File : VN082799.D
Acq On : 03 Jul 2024 13:38
Operator : JC\MD
Sample : PB161888ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161888ZHE#02

Quant Time: Jul 03 23:58:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

