

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070324\
 Data File : VN082809.D
 Acq On : 03 Jul 2024 17:40
 Operator : JC\MD
 Sample : P3152-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 446

Quant Time: Jul 04 00:02:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 28 06:02:21 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	114265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	217284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	216545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	88399	50.166	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.340%		
35) Dibromofluoromethane	8.171	113	72238	50.495	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.000%		
50) Toluene-d8	10.571	98	276353	51.654	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.300%		
62) 4-Bromofluorobenzene	12.853	95	102214	55.632	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 111.260%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	40249	60.286	ug/l	100
20) Methylene Chloride	5.271	84	5621	3.327	ug/l #	69
43) Isopropyl Acetate	8.694	43	143311	37.120	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070324\
Data File : VN082809.D
Acq On : 03 Jul 2024 17:40
Operator : JC\MD
Sample : P3152-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
446

Quant Time: Jul 04 00:02:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062724W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 28 06:02:21 2024
Response via : Initial Calibration

