

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070419\
 Data File : VN056596.D
 Acq On : 3 Jul 2019 19:48
 Operator : JC/SP
 Sample : K3680-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Jul 03 23:30:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	276510	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	478749	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	486869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	189856	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	161851	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
35) Dibromofluoromethane	7.59	113	139766	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
50) Toluene-d8	10.09	98	601101	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	12.40	95	215852	55.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.44%	
Target Compounds						
16) Acetone	3.82	43	3071	2.000	ug/l	98
19) Methyl tert-butyl Ether	5.05	73	309542	38.297	ug/l	98
29) Tetrahydrofuran	7.22	42	4229	3.405	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070419\
Data File : VN056596.D
Acq On : 3 Jul 2019 19:48
Operator : JC/SP
Sample : K3680-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Jul 03 23:30:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

